# AMENDMENT NO. 1 DECEMBER 2016
# TO
# IS 8432 : 1977 SPECIFICATION FOR TUBES, TRACHEOTOMY (RUBBER OR PLASTICS)

(*Page* 1, *clause* **3.1**, *sentence* 3) **—** Substitute the following for the existing sentence:

"Where the end is open, it shall have a female conical shape and capable of fitting to the male conical fitting with a 6 percent (LUER) taper in accordance with IS 3234 (Part 2): 1995 'Conical fittings with a 6% (LUER) taper for syringes, needles and certain other medical equipment: Part 2 Lock fitting (*second revision*)'."

(MHD 11)

Publication Unit, BIS, New Delhi, India